

EV6421J/EV6421M Digital thermoregulators for general purposes

GB ENGLISH

1 GETTING STARTED


1.1 Important

Read these instructions carefully before installing and using the instrument and follow all additional information for installation and electrical connection; keep these instructions close to the instrument for future consultations.

The instrument must be disposed according to the local legislation about the collection for electrical and electronic equipment.


1.2 Size

3 DIN modules; size in mm (in).



1.3 Installation


On DIN rail.



Additional information for installation:

- working conditions (working temperature, humidity, etc.) must be between the limits indicated in the technical data
- do not install the instrument close to heating sources (heaters, hot air ducts, etc.), devices provided with big magnetos (big speakers, etc.), locations subject to direct sunlight, rain, humidity, dust, mechanical vibrations or bumps
- according to the safety legislation, the protection against electrical parts must be ensured by a correct installation of the instrument; the parts that ensure the protection must be installed so that you can not remove them if not by using a tool.

1.4 Wiring diagram



With reference to the wiring diagram:

- the serial port (by request) is the port for the communication with the supervision system (through a serial interface, via TTL, with MODBUS communication protocol) or with the programming key; **the port must not be used at the same time for the same purposes**.

Additional information for electrical connection:

- do not operate on the terminal blocks with electrical or pneumatic screws
- if the instrument has been moved from a cold location to a warm one, the humidity could condense on the inside; wait about an hour before supplying it
- test the working power supply voltage, working electrical frequency and working electrical power of the instrument; they must correspond with the local power supply
- disconnect the local power supply before servicing the instrument
- provide the thermocouple with a protection able to protect it against contacts with metal parts or use insulated thermocouples
- do not use the instrument as safety device
- for repairs and information on the instrument please contact Evco sales network.

2 USER INTERFACE

2.1 Turning on/off the instrument

To turn on the instrument you have to supply it; to turn it off it is enough to cut off the power supply.

2.2 The display

If the instrument is turned on, during the normal operation the display will show the quantity you have set with parameter P5:

- if P5 = 0, the display will show the room temperature
- if P5 = 1, the display will show the working setpoint.

2.3 Showing the room temperature

- make sure the keyboard is not locked and no procedure is running
- press **set** 2 s: the display will show "Pb1"
- press **set**

To quit the procedure:

- press **set** or do not operate 60 s
- press **set** or **▼** as long as the display shows the quantity you have set with parameter P5 or do not operate 60 s.

2.4 Activating the defrost by hand

- make sure the keyboard is not locked and no procedure is running
- press **set** 4 s.

If parameter r5 has value 1 (heating action), the defrost functions will not be enabled.

2.5 Locking/unlocking the keyboard

To lock the keyboard:

- make sure no procedure is running
- press **set** and **▼** 2 s: the display will show "Loc" 1 s.

If the keyboard is locked, you will not be allowed to:

- activate the defrost by hand
- modify the working setpoint with the procedure related in paragraph 4.1 (you also can modify the working setpoint through parameter SP).

These operations provoke the visualization of the label "Loc" 1 s.

To unlock the keyboard:

- press **set** and **▼** 2 s: the display will show "UnL" 1 s.

2.6 Silencing the buzzer

- make sure no procedure is running
- press a button (the first pressure of the button does not provoke its usual effect).


3 OPERATION

3.1 Preliminary information

The operation mainly depends on parameter r5.


3.2 Operation with parameter r5 = 0 (cooling action)

temp.



3.3 Operation with parameter r5 = 1 (heating action)

temp.



When the cause that has provoked the alarm disappears, the instrument restores the normal operation.

7 INTERNAL DIAGNOSTICS

7.1 Internal diagnostics

7.2 Settings

4.1 Setting the working setpoint

- make sure the keyboard is not locked and no procedure is running
- press **set** LED **out 1** will flash
- press **▲** or **▼** in 15 s; also look at parameters r1, r2 and r3
- press **set** or do not operate 15 s.

You also can modify the working setpoint through parameter SP.

4.2 Setting configuration parameters

To gain access the procedure:

- make sure no procedure is running
- press **▲** and **▼** 4 s: the display will show "PA"
- press **set**
- press **▲** or **▼** in 15 s to set "-19"
- press **set** or do not operate 15 s
- press **▲** and **▼** 4 s: the display will show "SP".

To select a parameter:

- press **▲** or **▼**
- To modify a parameter:

- press **set**
- press **▲** or **▼** in 15
- press **set** or do not operate 15 s.

To quit the procedure:

- press **▲** and **▼** 4 s or do not operate 60 s.
- **Switch off/on the power supply of the instrument after the modification of the parameters.**

4.3 Restoring the default value of configurat. parameters

To make sure no procedure is running

- press **▲** and **▼** 4 s: the display will show "PA"
- press **set**
- press **▲** or **▼** in 15 s to set "743"
- press **set** or do not operate 15 s
- press **▲** and **▼** 4 s: the display will show "dEF"
- press **set**
- press **▲** or **▼** in 15 s to set "149"
- press **set** or do not operate 15 s
- switch off/on the power supply of the instrument.

4.4 Make sure the default value of the parameters is appropriate, in particular if the probes are not PTC probes.

5 SIGNALS

5.1 Signals

CODE

MEANING

LED

out 1

LED load

if it is lit, the load will be turned on

if it flashes:

- the modification of the working setpoint will be running
- a load protection will be running (parameters C1 and C2)

LED defrost

if it is lit, the defrost will be running

LED alarm

if it is lit, an alarm will be running

°C

LED Celsius degree

if it is lit, the unit of measure of the temperatures will be

Celsius degree (parameter P2)

°F

LED Fahrenheit degree

if it is lit, the unit of measure of the temperatures will be

Fahrenheit degree (parameter P2)

CODE

MEANING

Loc

the keyboard and/or the working setpoint are locked (parameter r3); also look at paragraph 2.5

6 ALARMS

6.1 Alarms

CODE

MEANING

AL1

First temperature alarm

Remedies:

- check the room temperature
- look at parameters A1 and A3

Effects:

• no effect

AL2

Second temperature alarm

Remedies:

- check the room temperature
- look at parameters A5 and A7

Effects:

- if parameter i5 has value 1, there will be no effect
- if parameter i5 has value 2, the load will be turned off

When the cause that has provoked the alarm disappears, the instrument restores the normal operation.

7.2 Internal diagnostics

CODE

MEANING

Pr1

Room probe error

Remedies:

- look at parameter P0
- check the integrity of the probe
- check the connection instrument-probe
- check the room temperature

Effects:

- the load activity will depend on parameters C4 and C5

When the cause that has provoked the alarm disappears, the instrument restores the normal operation.

8 TECHNICAL DATA

8.1 Technical data

Box: self-extinguishing grey.

Frontal protection: IP 54.

Connections: screw terminal blocks (power supply, inputs and outputs), 6 poles connector (serial port; by request).

Working temperature: from 0 to 55 °C (32 to 131 °F 10 ... 90% of relative humidity without condensate).

Power supply: 230 VAC, 50/60 Hz, 3 VA (approximate); 115 VAC or 24 VAC or 12-24 VAC/DC or 12 VAC/DC by request.

Alarm buzzer: by request.

Measure Inputs EV6421J: 1 (room probe) for J/K thermocouples.

Measure Inputs EV6421M: 1 (room probe) for PTC/NTC probes, J/K thermocouples, 2/3 wires Pt 100, Pt 1000 and Ni 120 probes, 0-20/40 mA and 0-10/20 V transducers (universal measure input).

9 WORKING SETPOINTS AND CONFIGURATION PARAMETERS

9.1 Working setpoints

	MIN.	MAX.	U.M.	DEF	WORKING SETPOINTS
r1	r2	°C/F (1)	0.0		working setpoint

9.2 Configuration parameters

PARAM	MIN.	MAX.	U.M.	DEF	WORKING SETPOINTS
SP	r1	r2	°C/F (1)	0.0	working setpoint

PARAM	MIN.	MAX.	U.M.	DEF	MEASURE INPUTS (2)
CA1	-25.0	25.0	°C/F (1)	0.0	room probe offset

P0	0	13	---	5	kind of probe 0 = PTC 1 = NTC 2 = J 3 = K 4 = 3 wires Pt 100 5 = 2 wires Pt 100 6 = 3 wires Pt 1000 7 = 2 wires Pt 1000 8 = 4-20 mA 9 = 0-20 mA 10 = 2-10 V 11 = 0-10 V 12 = 3 wires Ni 120 13 = 2 wires Ni 120
----	---	----	-----	---	---

P1	0	1	---	1	if P0 = 0 ... 7 or 12 ... 13, decimal point Celsius degree 1 = YES if P0 = 8 ... 11, decimal point position 0 = no decimal point 1 = on the digit of ten
----	---	---	-----	---	--

P2	0	2	---	0	unit of measure temperature [influential only on LED Celsius degree and on LED Fahrenheit if P0 = 8 ... 11] (3) (4) 0 = °C 1 = °F 2 = LED Celsius degree and LED Fahrenheit degree will remain turned off
----	---	---	-----	---	--

P3	-199.0	199.0	points	-20.0	minimum value of the range of the transducer
----	--------	-------	--------	-------	--

P4	-199.0	199.0	points	80.0	maximum value of the range of the transducer
----	--------	-------	--------	------	--

P5	0	1	---	0	quantity to show during the normal operation
----	---	---	-----	---	--

PARAM	MIN.	MAX.	U.M.	DEF	MAIN REGULATOR
-------	------	------	------	-----	----------------

r0	0.1	99.0	°C/F (1)	2.0	working setpoint differential
----	-----	------	----------	-----	-------------------------------

r1	-199.0	r2	°C/F (1)	0.0	minimum working setpoint
----	--------	----	----------	-----	--------------------------

r2	r1	[5]	°C/F (1)	350.0	maximum working setpoint
----	----	-----	----------	-------	--------------------------

r3	0	1	---	0	locking the working setpoint modification [with the procedure related in paragraph 4.1] 1 = YES
----	---	---	-----	---	--

r4	-99.0	99.0	°C/F (1)	0.0	temperature variation during function Energy Saving; also look at i5
----	-------	------	----------	-----	--

r5	0	1	---	[6]	cooling or heating action 0 = cooling
----	---	---	-----	-----	--

PARAM	MIN.	MAX.	U.M.	DEF	LOAD PROTECTIONS
-------	------	------	------	-----	------------------

C1	0	240	min	0	minimum time between two activations in succession of the load; also load delay since the end of the room probe error (7)
----	---	-----	-----	---	---

C2	0	240	min	0	minimum time the load remains turned off; also load delay since you turn on the instrument
----	---	-----	-----	---	--

C3	0	240	s	0	minimum time the load remains turned on
----	---	-----	---	---	---

C4	0	240	min	10	time the load remains turned off during the room probe error; also look at C5
----	---	-----	-----	----	---

C5	0	240	min	10	time the load remains turned on during the room probe error; also look at C4
----	---	-----	-----	----	--

PARAM	MIN.	MAX.	U.M.	DEF	DEFROST (8)
-------	------	------	------	-----	-------------

d0	0	99	h	8	defrost interval (9) 0 = the defrost at intervals will never be activated
----	---	----	---	---	--

d3	0	99	min	0	defrost duration 0 = the defrost will never be activated
----	---	----	-----	---	---

d4	0	1	---	0	defrost when you turn on the instrument 1 = YES
----	---	---	-----	---	--

d5	0	99	min	0	defrost delay when you turn on the instrument (only if d4 = 1)
----	---	----	-----	---	--

d6	0	1	---	1	temperature shown during the defrost 0 = room temperature 1 = if to the defrost activation the room temperature is below "working setpoint + r0"; at most "working setpoint + r0"; if to the defrost activation the room temperature is above "working setpoint + r0"; at most the room temperature to the defrost activation (10)
----	---	---	-----	---	--

PARAM	MIN.	MAX.	U.M.	DEF	TEMPERATURE ALARMS
-------	------	------	------	-----	--------------------

A1	-199.0	[5]	°C/F (1)	0.0	temperature the first temperature alarm is activated; also look at A3 (11)
----	--------	-----	----------	-----	--

A2	0	240	min	0	first temperature alarm delay (12)
----	---	-----	-----	---	------------------------------------

A3	0	4	---	0	kind of first temperature alarm 0 = alarm not enabled 1 = absolute lower alarm (or A1) 2 = absolute upper alarm (or A1) 3 = lower alarm relative to the working setpoint (or "working setpoint - A1"; consider A1 without sign, do not consider r4)
----	---	---	-----	---	---

ITALIANO

9 SETPOINT DI LAVORO E PARAMETRI DI CONFIGURAZIONE

9.1 Setpoint di lavoro

SETPOINT DI LAVORO

setpoint di lavoro

9.2 Parametri di configurazione