



EVK403J/EVK403M Two outputs digital thermoregulators (with alarm relay) for general purposes

ENGLISH**1 GETTING STARTED****1.1 Important**


Read these instructions carefully before installing and using the instrument and follow all additional information for installation and electrical connection; keep these instructions close to the instrument for future consultations.

1.2 Installing the instrument

Panel mounting, with click brackets (supplied by the builder); dimensions in mm (in).



Additional information for installation:
 • 73.0 (2.874) is the maximum depth with screw terminal blocks
 • 83.0 (3.267) is the maximum depth with extractable terminal blocks
 • the panel thickness must not be higher than 8.0 mm (0.314 in)
 • working conditions (working temperature, humidity, etc.) must be between the limits indicated in the technical data
 • do not install the instrument close to heating sources (heaters, hot air ducts, etc.), devices provided with big magnetics (big speakers, etc.), locations subject to direct sunlight, rain, humidity, dust, mechanical vibrations or bumps
 • according to the safety legislation, the protection against electrical parts must be ensured by a correct installation of the instrument; the parts that ensure the protection must be installed so that you can not remove them if not by using a tool.

1.3 Wiring diagram

With reference to the wiring diagram:
 • terminals 4 and 5 are available only in the models with power supply 230 VAC and 115 VAC; terminals 6 and 7 are available only in the models with power supply 12 VAC/DC and 12-24 VAC/DC

- the serial port (by request) is the port for the communication with the supervision system (through a serial interface, via TTL, with MODBUS communication protocol) or with the programming key; the port must not be used at the same time for the same purposes.

Additional information for electrical connection:

- do not operate on the terminal blocks with electrical or pneumatic screwdrivers
- do not operate on the terminal blocks with electrical or pneumatic screwdrivers
- if the instrument has been moved from a cold location to a warm one, the humidity could condense on the inside; wait about an hour before supplying it
- test the working power supply voltage, working electrical frequency and working electrical power of the instrument; they must correspond with the local power supply
- disconnect the local power supply before servicing the instrument
- provide the thermocouple with a protection able to protect it against contacts with metal parts or use insulated thermocouples
- do not use the instrument as safety device
- for repairs and information on the instrument please contact Evco sales network.

2 USER INTERFACE**2.1 Turning on/off the instrument**

To turn on the instrument you have to supply it; to turn it off it is enough to cut off the power supply.

2.2 The display

If the instrument is turned on, during the normal operation the display will show the quantity you have set with parameter P5:

- if P5 = 0, the display will show the room temperature
- if P5 = 1, the display will show the first working setpoint.

2.3 Showing the room temperature

- make sure the keyboard is not locked and no procedure is running
- press **set** 2 s: the display will show "Pb1"

To quit the procedure:

- press **set** or do not operate 60 s
- press **▲** or **▼** as long as the display shows the quantity you have set with parameter P5 or do not operate 60 s.

2.4 Locking/unlocking the keyboard

To lock the keyboard:

- make sure no procedure is running
- press **set** and **▼** 2 s: the display will show "Loc" 1 s.

If the keyboard is locked, you will not be allowed to:

- modify the working setpoints with the procedures related in paragraphs 4.1 and 4.2 (you also can modify the working setpoints through parameters SP1 and SP2).

This operation provokes the visualization of the label "Loc" 1 s.

To unlock the keyboard:

- press **set** and **▼** 2 s: the display will show "Unl" 1 s.

2.5 Silencing the buzzer

- make sure no procedure is running
- press a button (the first pressure of the button does not provoke its usual effect).

3 OPERATION**3.1 Preliminary information**

The operation mainly depends on parameter CFG.

3.2 Operation with parameter CFG = 1 (the first working setpoint is independent and the second one is relative to the first)

If parameter CFG has value 4, the second working setpoint will not be available and parameters SP2, r5, r6, r7, r8, r9 and r10 will not be significant.

You can get each load to work for cooling (parameter r5 = 0) or for heating (parameter r5 = 1); parameter r5 sets the action for each load. In this example each load works for cooling.

4 SETTINGS**4.1 Setting the first working setpoint**

- make sure the keyboard is not locked and no procedure is running
- press **set** LED **out 1** will flash
- press **▲** or **▼** in 15 s; also look at parameters r1, r2 and r3
- do not operate 15 s.

You also can modify the first working setpoint through parameter SP1.

4.2 Setting the second working setpoint

- press **set** during the modification of the first working setpoint: LED **out 2** will flash
- press **▲** or **▼** in 15 s; also look at parameters r7, r8 and r9
- press **set** or do not operate 15 s.

You also can modify the second working setpoint through parameter SP2.


If parameter CFG has value 1, you can set the second working setpoint through parameter SP2 only (because it is relative to the first one). You can get each load to work for cooling (parameters r5 and r10 = 0) or for heating (parameters r5 and r10 = 1).

In this example load 1 works for cooling, load 2 works for heating and the second working setpoint has a negative value.

4.3 Setting configuration parameters

To gain access the procedure:

- make sure no procedure is running
- press **set** and **▼** 4 s: the display will show "PA"
- press **set**
- press **▲** or **▼** in 15 s to set "-19"
- press **set** or do not operate 15 s
- press **set** and **▼** 4 s: the display will show "SP1".

3.3 Operation with parameter CFG = 2 (two independent working setpoints)

To select a parameter:

- press **▲** or **▼**
- To modify a parameter:
- press **set**
- press **▲** or **▼** in 15 s
- press **set** or do not operate 15 s.









To quit the procedure:

- press **▲** and **▼** 4 s or do not operate 60 s.

Switch off/on the power supply of the instrument after the modification of the parameters.**4.4 Restoring the default value of configuration parameters**

- make sure no procedure is running
- press **set** and **▼** 4 s: the display will show "PA"
- press **set**
- press **▲** or **▼** in 15 s to set "743"
- press **set** or do not operate 15 s
- press **set** and **▼** 4 s: the display will show "DEF"
- press **set**
- press **▲** or **▼** in 15 s to set "149"
- press **set** or do not operate 15 s: the display will show "DEF" flashing 4 s, after which the instrument will quit the procedure
- switch off/on the power supply of the instrument.

Make sure the default value of the parameters is appropriate, in particular if the probes are not J thermocouples.

5 SIGNALS**5.1 Signals****6 ALARMS****6.1 Alarms**

When the cause that has provoked the alarm disappears, the instrument restores the normal operation.

8 TECHNICAL DATA**8.1 Technical data**

Box: self-extinguishing grey.

Frontal protection: IP 65.

Connections (use copper conductors only): screw terminal blocks (power supply, inputs and outputs), 6 poles connector (serial port; by request); spring extractable terminal blocks (power supply, inputs and outputs) by request.

Working temperature: from 0 to 55 °C (32 to 131 °F, 10 ... 90% of relative humidity without condensate).

Power supply: 230 VAC, 50/60 Hz, 3 VA (approximate); 115 VAC or 12-24 VAC/DC or 12 VAC/DC by request.

Insulation class: 2.

Alarm buzzer: by request.

Measure inputs EVK403J: 1 (room probe) for J/K thermocouples.

Measure inputs EVK403M: 1 (room probe) for PTC/NTC probes, J/K thermocouples, 2/3 wires Pt 100, Pt 1000 and Ni 120 probes, 0-20/4-20 mA and 0-10/2-10 V transducers (universal measure input).

Digital inputs: 1 (multipurpose) for NO/NC contact (free of voltage, 5 V 1 mA).

Working range: from -50 to 150 °C (-50 to 300 °F) for PTC probe, from -40 to 110 °C (-40 to 230 °F) for NTC probe, from -100 to 800 °C (-140 to 1,450 °F) for J thermocouple, from -100 to 1,300 °C (-140 to 1,999 °F) for K thermocouple, from -200 to 650 °C (-320 to 1,200 °F) for 2/3 wires Pt 100 probe, from -80 to 300 °C (-110 to 570 °F) for 2/3 wires Pt 1000 probe, from -80 to 300 °C (-110 to 570 °F) for 2/3 wires Ni 120 probe.

Resolution: 0.1 °C/1 °C/1 °F.

Digital outputs: 3 relays:

- Load 1 relay:** 16 res. A @ 250 VAC, 5 FLA, 30 LRA (change-over contact)
- Load 2 relay:** 8 res. A @ 250 VAC, 2 FLA, 12 LRA (change-over contact)
- alarm relay:** 8 res. A @ 250 VAC, 2 FLA, 12 LRA (change-over contact)

3.4 Operation with parameter CFG = 3 (neutral zone)</

Se il parametro CFG è impostato a 3 o 4, il secondo setpoint di lavoro non sarà disponibile.

4.3 Impostazione dei parametri di configurazione

Per accedere alla procedura:

- assicurarsi che non sia in corso alcuna procedura
- premere **▲** e **▼** per 4 s: il display visualizzerà "PA"
- premere **set**
- premere **▲** o **▼** entro 15 s per impostare "-19"
- premere **set** o non operare per 15 s
- premere **▲** e **▼** per 4 s: il display visualizzerà "SP1".

Per selezionare un parametro:

- premere **▲** o **▼**

Per modificare un parametro:

- premere **set**
- premere **▲** o **▼** entro 15 s
- premere **set** o non operare per 15 s.

Per uscire dalla procedura:

- premere **▲** e **▼** per 4 s o non operare per 60 s.

Interrompere l'alimentazione dello strumento dopo la modifica dei parametri.

4.4 Ripristino del valore di default dei parametri di configurazione

- assicurarsi che non sia in corso alcuna procedura
- premere **▲** e **▼** per 4 s: il display visualizzerà "PA"
- premere **set**

- premere **▲** o **▼** entro 15 s per impostare "743"
- premere **set** o non operare per 15 s

- premere **▲** e **▼** per 4 s: il display visualizzerà "dEF"
- premere **set**

- premere **▲** o **▼** entro 15 s per impostare "149"
- premere **set** o non operare per 15 s: il display visualizzerà "dEF" lampeggiante per 4 s, dopodiché lo strumento uscirà dalla procedura

• interrompere l'alimentazione dello strumento.

Accertarsi che il valore di default dei parametri sia oppure, in particolare se le sonde non sono termocopie J.

5 SEGNALAZIONI

5.1 Segnalazioni

LED	SIGNIFICATO
-----	-------------

out 1	LED carico 1 se è acceso, il carico 1 sarà acceso se lampeggia: • sarà in corso la modifica del primo setpoint di lavoro • sarà in corso una protezione del carico 1 (parametri C1 e C2)
--------------	--

out 2	LED carico 2 se è acceso, il carico 2 sarà acceso se lampeggia: • sarà in corso la modifica del secondo setpoint di lavoro • sarà in corso una protezione del carico 2 (parametri C7 e C8)
--------------	--

A	LED allarme se è acceso, sarà in corso un allarme
----------	--

°C	LED grado Celsius se è acceso, l'unità di misura delle temperature sarà il grado Celsius (parametro P2)
-----------	--

°F	LED grado Fahrenheit se è acceso, l'unità di misura delle temperature sarà il grado Fahrenheit (parametro P2)
-----------	--

CODICE	SIGNIFICATO
---------------	--------------------

Loc	la tastiera e/o i setpoint di lavoro sono bloccati (parametri r3 e/o r9); si veda il paragrafo 2.4
------------	--

6 ALLARMI

6.1 Allarmi

CODICE	SIGNIFICATO
--------	-------------

AL1	Primo allarme di temperatura Rimedi: • verificare la temperatura dell'ambiente • si vedano i parametri A1 e A3
------------	---

Conseguenze:	• l'uscita di allarme verrà accesa
---------------------	------------------------------------

AL2 Secondo allarme di temperatura

Rimedi:
• verificare la temperatura dell'ambiente
• si vedano i parametri A5 e A7

Conseguenze:

- l'uscita di allarme verrà accesa

iA Allarme ingresso multifunzione

Rimedi:
• verificare le cause che hanno provocato l'attivazione dell'ingresso

- si vedano i parametri i1 e i5

Conseguenze:

- if P0 = 0 ... 7 or 12 ... 13, decimal point Celsius

P1 0 1 --- 1 if P0 = 0 ... 7 or 12 ... 13, punto decimale grado Celsius

1 = YES

if P0 = 8 ... 11, decimal point position

0 = no decimal point

1 = on the digit of ten

se P0 = 8 ... 11, posizione del punto decimale

0 = nessun punto decimale

1 = sul digit delle decine

unità di misura temperatura (influenza solo sul LED

Celsius degree and on LED grado Fahrenheit se se

P0 = 8 ... 11) (3) (4)

0 = °C

1 = °F

2 = LED Celsius degree and LED grado Fahrenheit degree

rimarranno spenti

valore minimo della taratura del trasduttore

valore massimo della taratura del trasduttore

grandezza visualizzata durante il normale funziona-

mento

0 = temperatura dell'ambiente

1 = primo setpoint di lavoro

REGOLATORI

r0 0.1 99.0 °C/F (1) 2.0 if CFG = 1 o 2, differenziale del primo setpoint di

lavoro

if CFG = 3, differenziale del carico funzionante per

freddo (carico 1)

minimo primo setpoint di lavoro

massimo primo setpoint di lavoro

blocco della modifica del primo setpoint di lavoro

(con la procedura indicata nel paragrafo 4.1)

1 = YES

locking the first working setpoint modification (with

the procedure related in paragraph 4.1)

se CFG = 1 o 2, funzionamento per freddo o per

caldo del carico 1

se CFG = 4, funzionamento per freddo o per caldo

dei carichi

0 = per freddo

se CFG = 1 o 2, differenziale del secondo setpoint di

lavoro

if CFG = 3, differenziale del carico funzionante per

caldo (carico 2)

minimo secondo setpoint di lavoro

massimo secondo setpoint di lavoro

blocco della modifica del secondo setpoint di lavoro

(con la procedura indicata nel paragrafo 4.2)

1 = SI

functionamento per freddo o per caldo del carico 2

(solo se CFG = 1 o 2)

0 = per freddo

se CFG = 3, valore della zona neutra

se CFG = 4, valore di due gradini

PROTEZIONI DEI CARICHI

C1 0 240 min 0 minimum time between two activations in suc-

cession of load 1; also load 1 delay since the end of the

room probe error [6]

C2 0 240 min 0 minimum time load 1 remains turned off; also

load 1 delay since you turn on the instrument

durata minima dello spegnimento del carico 1; an-

che ritardo carico 1 dall'accensione dello strumento

durata minima dell'accensione del carico 1

attività del carico 1 durante l'errore sonda ambiente

0 = spento

1 = acceso

C7 0 240 min 0 tempo minimo tra due accensioni consecutive del

carico 2; anche ritardo carico 2 dalla conclusione

dell'errore sonda ambiente [6]

C8 0 240 min 0 durata minima dello spegnimento del carico 2;

anche ritardo carico 2 dall'accensione dello strumento

durata minima dell'accensione del carico 2

attività del carico 2 durante l'errore sonda ambiente

0 = spento

1 = acceso

ALLARMI DI TEMPERATURA

temperatura alla quale viene attivato il primo allar-

me di temperatura; si veda anche A3 [7]

A1 -199.0 [5] °C/F (1) 0.0 temperature the first temperature alarm is activated;

also look at A3 [7]

A2 0 240 min 0 first temperature alarm delay

ritardo primo allarme di temperatura

A3 0 4 --- 0 kind of first temperature alarm

0 = alarm not enabled

1 = absolute lower alarm (or A1)

2 = absolute upper alarm (or A1)

3 = lower alarm relative to the first working setpoint