

# EVKB22 and EVKB32 Simple digital thermostats for normal and low temperature refrigerating units

## GB ENGLISH

### 1 GETTING STARTED



#### 1.1 Important

Read these instructions carefully before installing and using the instrument and follow all additional information for installation and electrical connection; keep these instructions close to the instrument for future consultations.

The instrument must be disposed according to the local legislation about the collection for electrical and electronic equipment.

#### 1.2 Installing the instrument

Panel mounting, with click brackets (supplied by the builder); dimensions in mm (in).




Additional information for installation:

- 59.0 (2.322) is the maximum depth with screw terminal blocks
- 81.5 (3.208) is the maximum depth with extractable terminal blocks
- the panel thickness must not be higher than 8.0 mm (0.314 in)
- working conditions (working temperature, humidity, etc.) must be between the limits indicated in the technical data
- do not install the instrument close to heating sources (heaters, hot air ducts, etc.), devices provided with big magnetics (big speakers, etc.), locations subject to direct sunlight, rain, humidity, dust, mechanical vibrations or bumps
- according to the safety legislation, the protection against electrical parts must be ensured by a correct installation of the instrument; the parts that ensure the protection must be installed so that you can not remove them if not by using a tool.


#### 1.3 Wiring diagram

With reference to the wiring diagrams:


- the load managed by relay K2 depends on parameter u0
- the serial port (by request) is the port for the communication with the programming key.




Models with power supply 230 VAC and 115 VAC.  
Modelli con alimentazione 230 VCA e 115 VCA.



Models with power supply 12 VAC/DC.  
Modelli con alimentazione 12 VCA/CC.



Models with power supply 230 VAC and 115 VAC.  
Modelli con alimentazione 230 VCA e 115 VCA.



Models with power supply 12 VAC/DC.  
Modelli con alimentazione 12 VCA/CC.

Additional information for electrical connection:

- do not operate on the terminal blocks with electrical or pneumatic screws
- if the instrument has been moved from a cold location to a warm one, the humidity could condense on the inside; wait about an hour before supplying it
- test the working power supply voltage, working electrical frequency and working electrical power of the instrument; they must correspond with the local power supply
- disconnect the local power supply before servicing the instrument
- do not use the instrument as safety device
- for repairs and information on the instrument please contact Evco sales network.

#### 2 USER INTERFACE

##### 2.1 Preliminary information

There are the following operation status:

- status "on" (the instrument is supplied and is turned on: the regulators can be turned on)
- status "stand-by" (the instrument is supplied but it is turned off via software: the regulators are turned off)
- status "off" (the instrument is not supplied).

"Turning on" means moving from status stand-by to status on; "turning off" means moving from status on to status stand-by.

After an interruption of power supply the instrument moves to the status it was before the interruption.

The load managed by relay K2 depends on parameter u0, as follows:

- if parameter u0 has value 0 (default value), the load will be the defrost system (a heater for electric defrost or a valve for hot gas defrost)
- if parameter u0 has value 1, the load will be the evaporator fan (the defrost occurs stopping the compressor).

##### 2.2 Turning on/off the instrument

• make sure the keyboard is not locked and no procedure is running  
• press **set** 4 s.

##### 2.3 The display

If the instrument is turned on, during the normal operation the display will show the cabinet temperature.

If the instrument is turned off, the display will be switched off.

##### 2.4 Showing the evaporator temperature

• make sure the keyboard is not locked and no procedure is running  
• press **set** 2 s: the display will show "P2"

To quit the procedure:

• press **set** or do not operate 60 s  
• press **set** or **▼** as long as the display shows the cabinet temperature or do not operate 60 s.

If the evaporator probe is not enabled (parameter P3 = 0), the label "P2" will not be shown.

##### 2.5 Activating the defrost by hand

• make sure the keyboard is not locked and no procedure is running  
• press **set** 4 s.

If the function of the evaporator probe is the one of defrost probe (parameter P3 = 1) and to the defrost activation the evaporator temperature is above the one you have set with parameter d2, the defrost will not be activated.

##### 2.6 Locking/unlocking the keyboard

To lock the keyboard:

- make sure no procedure is running
- press **set** and **▼** 1 s: the display will show "Lo" 1 s.

If the keyboard is locked, you will not be allowed to:

- turn on/off the instrument through button **set**
- show the evaporator temperature
- activate the defrost by hand
- modify the working setpoint with the procedure related in paragraph 3.1 (you also can modify the working setpoint through parameter SP). These operations provoke the visualization of the label "Lo" 1 s.

To unlock the keyboard:

- press **set** and **▼** 1 s: the display will show "Un" 1 s.

### 3 SETTINGS

#### 3.1 Setting the working setpoint

- make sure the keyboard is not locked and no procedure is running
- press **set**: LED **Lo** will flash
- press **▲** or **▼** in 15 s; also look at parameters r1 and r2
- press **set** or do not operate 15 s.

You also can modify the working setpoint through parameter SP.

#### 3.2 Setting configuration parameters

- To gain access the procedure (for the models without access password):
- make sure the keyboard is not locked and no procedure is running
  - press **▲** and **▼** 4 s: the display will show "SP".

To gain access the procedure (for the models with access password):

- make sure no procedure is running
- press **▲** and **▼** 4 s: the display will show "PA"
- press **set** or **▼** in 15 s to set "-19"
- press **set** or do not operate 15 s
- press **▲** and **▼** 4 s: the display will show "SP".

To select a parameter:

- press **▲** or **▼**

To modify a parameter:

- press **set**
- press **▲** or **▼** in 15 s
- press **set** or do not operate 15 s.

To quit the procedure:

- press **▲** and **▼** 4 s or do not operate 60 s.

#### Switch off/on the power supply of the instrument after the modification of the parameters.

### 4 SIGNALS

#### 4.1 Signals

LED	MEANING
	LED compressor if it is lit, the compressor will be turned on if it flashes: • the defrost will be required but a compressor protection will be running (parameters C0 and C2, only if parameter u0 has value 0) • the dripping will be running (parameter d7) • the heating of the freezing fluid will be running (parameter d4, only if parameter u0 has value 0)
	LED defrost if it is lit, the defrost will be running if it flashes: • the defrost will be required but a compressor protection will be running (parameters C0 and C2, only if parameter u0 has value 0) • the dripping will be running (parameter d7) • the heating of the freezing fluid will be running (parameter d4, only if parameter u0 has value 0)

LED	MEANING
	LED evaporator fan (only if parameter u0 has value 1) if it is lit, the evaporator fan will be turned on
	LED alarm if it is lit, an alarm or an error will be running
	LED Celsius degree if it is lit, the unit of measure of the temperatures will be Celsius degree (parameter P2)
	LED Fahrenheit degree if it is lit, the unit of measure of the temperatures will be Fahrenheit degree (parameter P2)

CODE	MEANING
<b>Lo</b>	the keyboard is locked; look at paragraph 2.6

#### 5 ALARMS

##### 5.1 Alarms

CODE	MEANING
<b>AL</b>	Lower temperature alarm Remedies: • check the cabinet temperature • look at parameter A1 Effects: • no effect
<b>AH</b>	Upper temperature alarm Remedies: • check the cabinet temperature • look at parameter A4 Effects: • no effect

CODE	MEANING
<b>IA</b>	Multipurpose input alarm (only if parameter i5 has value 1 or 2) Remedies: • check the reasons that have provoked the activation of the input • look at parameters i1 and i5 Effects: • if parameter i5 has value 1, there will be no effect • if parameter i5 has value 2, the compressor will be turned off

#### ITALIANO

### 1 PREPARATIVI

#### 1.1 Importante

Leggere attentamente queste istruzioni prima dell'installazione e prima dell'uso e seguire tutte le avvertenze per l'installazione e per il collegamento elettrico; conservare queste istruzioni con lo strumento per consultazioni future.

**Lo strumento deve essere smaltito secondo le normative locali in merito alla raccolta delle apparecchiature elettriche ed elettroniche.**

#### 1.2 Installazione

A pannello, con le staffe a scatto in dotazione (si vedano i disegni del paragrafo 1.2 della sezione in Inglese). Avvertenze per l'installazione:

- 59.0 è la profondità massima con morsettiera a vite
- 81.5 è la profondità massima con morsettiera estraibili
- lo spessore del pannello non deve essere superiore a 8.0 mm
- accertarsi che le condizioni di lavoro (temperatura di impiego, umidità, ecc.) rientrino nei limiti indicati nei dati tecnici
- non installare lo strumento in prossimità di fonti di calore (resistenze, condotti dell'aria calda, ecc.), di apparecchi con forti magneti (grossi diffusori, ecc.), di luoghi soggetti alla luce solare diretta, pioggia, umidità, polvere eccessiva, vibrazioni meccaniche o scosse

Queste operazioni provocano la visualizzazione della label "Lo" per 1 s. Per sbloccare la tastiera:

- premere **set** e **▼** per 1 s: il display visualizzerà "Un" per 1 s.

#### 3 IMPOSTAZIONI

##### 3.1 Impostazione del setpoint di lavoro

- assicurarsi che la tastiera non sia bloccata e che non sia in corso alcuna procedura

- premere **set**: il LED **Lo** lampeggerà

- premere **▲** o **▼** entro 15 s; si vedono anche i parametri r1 e r2.

È inoltre possibile impostare il setpoint di lavoro attraverso il parametro SP.

#### 3.2 Impostazione dei parametri di configurazione

Per accedere alla procedura (per i modelli senza password di accesso):

- assicurarsi che la tastiera non sia bloccata e che non sia in corso alcuna procedura

- premere **▲** e **▼** per 4 s: il display visualizzerà "SP".

Per accedere alla procedura (per i modelli con password di accesso):

- assicurarsi che non sia in corso alcuna procedura

- premere **▲** e **▼** per 4 s: il display visualizzerà "PA".

Per selezionare un parametro:

- premere **▲** o **▼**

Per modificare un parametro:

- premere **set</**

Conseguenze:  
 • se il parametro i5 è impostato a 3, il compressore e il ventilatore dell'evaporatore verranno spenti (quest'ultimo solo se il parametro u0 è impostato a 1)  
 • se il parametro i5 è impostato a 4, il ventilatore dell'evaporatore verrà spento (solo se il parametro u0 è impostato a 1)

Quando la causa che ha provocato l'allarme scompare, lo strumento ripristina il normale funzionamento.

## 6 DIAGNOSTICA INTERNA

### 6.1 Diagnistica interna

CODICE | SIGNIFICATO

**P1** Errore sonda cella

Rimedi:

- verificare il tipo di sonda
- verificare l'integrità della sonda
- verificare il collegamento strumento-sonda
- verificare la temperatura della cella

Conseguenze:

- il compressore verrà acceso

**P2** Errore sonda evaporatore

Rimedi:

- gli stessi del caso precedente ma relativamente alla sonda evaporatore

Conseguenze:

- se il parametro P3 è impostato a 1, lo sbrinamento durerà il tempo stabilito con il parametro d3
- se il parametro F0 è impostato a 1 o 2, lo strumento funzionerà come se il parametro fosse impostato a 0 (solo se il parametro u0 è impostato a 1)

Quando la causa che ha provocato l'allarme scompare, lo strumento ripristina il normale funzionamento.

## 7 DATI TECNICI

### 7.1 Dati tecnici

**Contenitore:** autoestinguente grigio.

**Grado di protezione del frontale:** IP 65.

**Connessioni:** morsettiere a vite (alimentazione, ingressi e uscite), connettore a 6 poli (porta seriale; su richiesta); morsettiere estraibili (alimentazione, ingressi e uscite) su richiesta.

**Temperatura di impiego:** da 0 a 55 °C (10 ... 90% di umidità relativa senza condensa).

**Alimentazione:** 230 VCA, 50/60 Hz, 3 VA (approssimativi); 115 VCA o 12 VCA/CC su richiesta.

**Ingressi di misura:** 2 (sonda cella e sonda evaporatore) per sonde PTC o NTC (a seconda del modello).

**Ingressi digitali:** 1 (multifunzione/micro porta) per contatto NA/NC (contatto pulito, 5 V 1 mA).

**Campo di misura:** da -50 a 150 °C per sonda PTC, da -40 a 105 °C per sonda NTC.

**Risoluzione:** 0,1 °C (tra -19,9 e 19,9 °C)/1 °C/1 °F.

**Uscite digitali:** 2 relè:

• **relè compressore:** 30 A res. @ 250 VCA nell'EVKB32; 16 A res. @ 250 VCA nell'EVKB22 con alimentazione 12 VCA/CC; 8 A res. @ 250 VCA altrimenti (contatto NA)

• **relè sbrinamento/ventilatore dell'evaporatore:** 8 A res. @ 250 VCA (contatto NA nell'EVKB32 con alimentazione 12 VCA/CC; contatto in scambio altrimenti).

**La corrente massima consentita sui carichi è di 10 A.**

**Porta seriale:** porta per la comunicazione con la chiave di programmazione; su richiesta.

## GB ENGLISH

### 8 WORKING SETPOINTS AND CONFIGURATION PARAMETERS

#### 8.1 Working setpoints

MIN.	MAX.	U.M.	DEF.	WORKING SETPOINTS
r1	r2	°C/F (1)	0.0	working setpoint

#### 8.2 Configuration parameters

PARAM.	MIN.	MAX.	U.M.	DEF.	WORKING SETPOINTS
SP	r1	r2	°C/F (1)	0.0	working setpoint
PARAM.	MIN.	MAX.	U.M.	DEF.	MEASURE INPUTS
o1	-25	25	°C/F (1)	0.0	cabinet probe offset
o2	-25	25	°C/F (1)	0.0	evaporator probe offset
P1	0	1	---	1	decimal point Celsius degree (for the quantity to show during the normal operation, between -9,9 and 19,9 °C) 1 = YES
P2	0	1	---	0	unit of measure temperature (2) 0 = °C 1 = °F
P3	0	2	---	1	evaporator probe function 0 = probe not enabled 1 = defrost probe and thermostat probe for the evaporator fan (meaningful for this last only if u0 = 1) 2 = thermostat probe for the evaporator fan (meaningful only if u0 = 1)
PARAM.	MIN.	MAX.	U.M.	DEF.	MAIN REGULATOR
r0	0.1	15.0	°C/F (1)	2.0	working setpoint differential
r1	-99	r2	°C/F (1)	(3)	minimum working setpoint
r2	r1	199	°C/F (1)	50	maximum working setpoint
PARAM.	MIN.	MAX.	U.M.	DEF.	COMPRESSOR PROTECTIONS
C0	0	199	min	0	compressor delay since you turn on the instrument
C2	0	199	min	3	minimum time the compressor remains turned off
C3	0	199	s	0	minimum time the compressor remains turned on
PARAM.	MIN.	MAX.	U.M.	DEF.	DEFROST
d0	0	99	h	8	defrost interval 0 = the defrost at intervals will never be activated
d1	0	1	---	0	kind of defrost (visible only if u0 = 0) 0 = electric defrost 1 = hot gas defrost
d2	-99	99	°C/F (1)	2.0	defrost cutoff temperature (meaningful only if P3 = 1)
d3	0	99	min	30	if P3 = 0 or 2, defrost duration if P3 = 1, defrost maximum duration 0 = the defrost will never be activated
d4	0	1	---	0	defrost when you turn on the instrument 1 = YES
d5	0	199	min	0	defrost delay when you turn on the instrument (meaningful only if d4 = 1)
d6	0	1	---	1	temperature shown during the defrost 0 = cabinet temperature 1 = if to the defrost activation the cabinet temperature is below "working setpoint + r0", at most "working setpoint + r0"; if to the defrost activation the cabinet temperature is above "working setpoint + r0", at most the cabinet temperature to the defrost activation (4)
d7	0	15	min	2	dripping duration

## ITALIANO

### 8 SETPOINT DI LAVORO E PARAMETRI DI CONFIGURAZIONE

#### 8.1 Setpoint di lavoro

##### SETPOINT DI LAVORO

setpoint di lavoro

#### 8.2 Parametri di configurazione

##### SETPOINT DI LAVORO

setpoint di lavoro

##### INGRESSI DI MISURA

offset sonda cella

offset sonda evaporatore

punto decimale grado Celsius (per la grandezza visualizzata durante il normale funzionamento, tra -9,9 e 19,9 °C)  
1 = SI

P1  
P2  
P3  
P4  
P5  
P6  
P7  
P8  
P9  
P10  
P11  
P12  
P13  
P14  
P15  
P16  
P17  
P18  
P19  
P20  
P21  
P22  
P23  
P24  
P25  
P26  
P27  
P28  
P29  
P30  
P31  
P32  
P33  
P34  
P35  
P36  
P37  
P38  
P39  
P40  
P41  
P42  
P43  
P44  
P45  
P46  
P47  
P48  
P49  
P50  
P51  
P52  
P53  
P54  
P55  
P56  
P57  
P58  
P59  
P60  
P61  
P62  
P63  
P64  
P65  
P66  
P67  
P68  
P69  
P70  
P71  
P72  
P73  
P74  
P75  
P76  
P77  
P78  
P79  
P80  
P81  
P82  
P83  
P84  
P85  
P86  
P87  
P88  
P89  
P90  
P91  
P92  
P93  
P94  
P95  
P96  
P97  
P98  
P99  
P100  
P101  
P102  
P103  
P104  
P105  
P106  
P107  
P108  
P109  
P110  
P111  
P112  
P113  
P114  
P115  
P116  
P117  
P118  
P119  
P120  
P121  
P122  
P123  
P124  
P125  
P126  
P127  
P128  
P129  
P130  
P131  
P132  
P133  
P134  
P135  
P136  
P137  
P138  
P139  
P140  
P141  
P142  
P143  
P144  
P145  
P146  
P147  
P148  
P149  
P150  
P151  
P152  
P153  
P154  
P155  
P156  
P157  
P158  
P159  
P160  
P161  
P162  
P163  
P164  
P165  
P166  
P167  
P168  
P169  
P170  
P171  
P172  
P173  
P174  
P175  
P176  
P177  
P178  
P179  
P180  
P181  
P182  
P183  
P184  
P185  
P186  
P187  
P188  
P189  
P190  
P191  
P192  
P193  
P194  
P195  
P196  
P197  
P198  
P199  
P200  
P201  
P202  
P203  
P204  
P205  
P206  
P207  
P208  
P209  
P210  
P211  
P212  
P213  
P214  
P215  
P216  
P217  
P218  
P219  
P220  
P221  
P222  
P223  
P224  
P225  
P226  
P227  
P228  
P229  
P230  
P231  
P232  
P233  
P234  
P235  
P236  
P237  
P238  
P239  
P240  
P241  
P242  
P243  
P244  
P245  
P246  
P247  
P248  
P249  
P250  
P251  
P252  
P253  
P254  
P255  
P256  
P257  
P258  
P259  
P260  
P261  
P262  
P263  
P264  
P265  
P266  
P267  
P268  
P269  
P270  
P271  
P272  
P273  
P274  
P275  
P276  
P277  
P278  
P279  
P280  
P281  
P282  
P283  
P284  
P285  
P286  
P287  
P288  
P289  
P290  
P291  
P292  
P293  
P294  
P295  
P296  
P297  
P298  
P299  
P300  
P301  
P302  
P303  
P304  
P305  
P306  
P307  
P308  
P309  
P310  
P311  
P312  
P313  
P314  
P315  
P316  
P317  
P318  
P319  
P320  
P321  
P322  
P323  
P324  
P325  
P326  
P327  
P328  
P329  
P330  
P331  
P332  
P333  
P334  
P335  
P336  
P337  
P338  
P339  
P340  
P341  
P342  
P343  
P344  
P345  
P346  
P347  
P348  
P349  
P350  
P351  
P352  
P353  
P354  
P355  
P356  
P357  
P358  
P359  
P360  
P361  
P362  
P363  
P364  
P365  
P366  
P367  
P368  
P369  
P370  
P371  
P372  
P373  
P374  
P375  
P376  
P377  
P378  
P379  
P380  
P381  
P382  
P383  
P384  
P385  
P386  
P387  
P388  
P389  
P390  
P391  
P392  
P393  
P394  
P395  
P396  
P397  
P398  
P399  
P400  
P401  
P402  
P403  
P404  
P405  
P406  
P407  
P408  
P409  
P410  
P411  
P412  
P413  
P414  
P415  
P416  
P417  
P418  
P419  
P420  
P421  
P422  
P423  
P424  
P425  
P426  
P427  
P428  
P429  
P430  
P431  
P432  
P433  
P434  
P435  
P436  
P437  
P438  
P439  
P440  
P441  
P442  
P443  
P444  
P445  
P446  
P447  
P448  
P449  
P450  
P451  
P452  
P453  
P454  
P455  
P456  
P457  
P458  
P459  
P460  
P461  
P462  
P463  
P464  
P465  
P466  
P467  
P468  
P469  
P470  
P471  
P472  
P473  
P474  
P475  
P476  
P477  
P478  
P479  
P480  
P481  
P482  
P483  
P484  
P485  
P486  
P487  
P488  
P489  
P490  
P491  
P492  
P493  
P494  
P495  
P496  
P497  
P498  
P499  
P500  
P501  
P502  
P503  
P504  
P505  
P506  
P507  
P508  
P509  
P510  
P511  
P512  
P513  
P514  
P515  
P516  
P517  
P518  
P519  
P520  
P521  
P522  
P523  
P524  
P525  
P526  
P527  
P528  
P529  
P530  
P531