

RK 800X/RK 801X Multifunction digital controller for electric ovens

ENGLISH


1 GETTING STARTED

1.1 Important
Read these instructions carefully before installing and using the instrument; do not forget following all additional information for installation and electrical connection.


Keep these instructions close to the instrument for future consultations.

1.2 Installing the instrument

Panel mounting, with screw brackets.



Dimensions in mm (in).




DIMENS.	MINIMUM	TYPICAL	MAXIMUM
A	92.0 (3.622)	92.0 (3.622)	92.8 (3.653)
B	92.0 (3.622)	92.0 (3.622)	92.8 (3.653)

Additional information for installation:

- the maximum panel thickness must be 4 mm (0.157 in)
- position the brackets as indicated; moderate the clamping torque, in order not to damage box and screw brackets
- working conditions (ambient temperature, humidity, etc.) must be between the limits indicated in the technical data
- install the instrument in locations with suitable ventilation, in order to avoid the overheating of the instrument
- do not install the instrument close to heating sources (resistances, hot air ducts, etc.), locations subject to direct sunlight, rain, humidity, dust, mechanical vibrations or bumps, devices provided with big magnetos (big speakers, etc.)
- according to safety norms, the protection against electrical parts must be ensured by a correct installation of the instrument; the parts that ensure the protection must be installed so that you can not remove them if not by using a tool.

1.3 Wiring diagram



Additional information for electrical connection:

- do not operate on the terminal blocks with electrical or pneumatic screwdriver
- if the instrument has been moved from a cold to a warm location, the humidity will condense on the inside; wait about an hour before applying power to the instrument
- test the working power supply voltage, working electrical frequency and working electrical power of the instrument; they must correspond with the local power supply
- disconnect the local power supply before servicing the instrument
- give the probes a protection able to protect them against contacts with metal parts or use insulated probes
- do not use the instrument as safety device
- do not try repairing the instrument yourself; for repairs, always use the sales network
- for any further information concerning the instrument, please consult Evco.

2 CONFIGURING THE INSTRUMENT

2.1 Preliminary information

You can choose the users for relays K1, K2 and K3 (among four combinations).

INST. CODE	RELAY K1	RELAY K2	RELAY K3
1	output for regulation	airhole output	alarm output
2	output for regulation	chamber light output	output for cooking timer
3	output for regulation	chamber light output	output for steam injection
4	output for regulation	airhole output	output for acoustic signalings

Further features for RK 800X:

- cooking timer.

Further features for RK 801X (in addition to the ones for RK 800X):

- real time clock and function Programmed starting.

2.2 Selecting the instrument code

To gain access the procedure:

- switch off the power supply of the instrument
- restore the power supply

press 3 times in 4 s since the power supply has been restored: the instrument will show "SEL" flashing in the display at the top.

Otherwise:

- press 1 s to turn the instrument off
- press and 4 s: the instrument will show the instrument code in the display at the top

- press or to select "PA"
- press or in 15 s to set "743"
- press or 4 s: the instrument will show "SEL" flashing in the display at the top.

To modify the instrument code:

- press
- press or in 15 s
- press

To quit the procedure:

- press 1 s or switch off the power of the instrument.

If you modify the instrument code, the instrument will not lose the value of configuration parameters.

2.3 Restoring default configuration parameters

- gain access the procedure to select the configuration
- press the instrument will show "dEF" flashing in the display at the top

- press
- press or in 15 s to set "149"
- press the instrument will show "SEL" flashing and the buzzer will utter a short beep.

To quit the procedure:

- press 1 s or switch off the power supply of the instrument.

3 USER INTERFACE

3.1 Preliminary information

If the instrument is turned on:

- the display at the top will show the chamber temperature or the working setpoint value
- the display at the bottom will show the count of the cooking timer (if the time will be running) or the real time (only RK 801X; the display will be switched off for RK 800X).

If the instrument is turned off:

- the display at the top will be switched off
- the display at the bottom will show the real time (only RK 801X, look at parameter c7 as well; the display will be switched off for RK 800X).

3.2 Turning the instrument on/off

- press 1 s.

To turn off means turning the instrument off via software (the instrument is connected with the power supply).

3.3 How to know which the quantity showed by the instrument in the display at the top is

- press the instrument will show 2 s an indication in the display at the bottom.

INDICAT.	MEANING
Sn1	Chamber temperature
SEt	Working setpoint value

3.4 Selecting the quantity to show in the display at the top during the normal operation

- make sure the instrument is turned on
- press 4 s.

Afterwards the instrument will show 2 s one of the indications indicated in paragraph 3.3 in the display at the bottom.

3.5 Silencing the alarms

- press a button.

This will also deactivate the output for acoustic signalings (if present).

3.6 Turning the airhole on/off by hand (if present)

- make sure the instrument is turned on
- press and 1 s.

look at parameters c5 and c6 as well.

3.7 Turning the chamber light on/off (if present)

- make sure the instrument is turned on

press

4 REAL TIME CLOCK (ONLY RK 801X)

4.1 Setting the clock

To gain access the procedure:

- press 1 s: the instrument will show "rtc" in the display at the top and the real time (hours:minutes) in the display at the bottom (the left part of the display will flash).

To modify the hours:

- press or in 15 s (1).

To modify the minutes:

- press during the modification of the hours, then ...

- press or in 15 s.

To quit the procedure:

- press during the modification of the minutes or do not operate 15 s.

- (1) during the modification the instrument will show "rtc" in the display at the top.

5 PROGRAMMED STARTING (ONLY RK 801X)

5.1 Setting the starting time

To gain access the procedure:

- make sure the instrument is turned on
- press the instrument will show "tin" in the display at the top and the cooking timer value (hours:minutes) in the display at the bottom (the left part of the display will flash)

- press in 15 s: the instrument will show "dEL" in the display at the top and the starting time (hours:minutes) in the display at the bottom (the left part of the display will flash).

- press or in 15 s to set "743"

- press or 4 s: the instrument will show "SEL" flashing in the display at the top.

- press or in 15 s.

- press during the modification of the hours, then ...

- press or in 15 s.

- press during the modification of the minutes: the instrument will show "int" in the display at the top and the number of days in the display at the bottom

- press or in 15 s (2).

- press during the modification of the days or do not operate 15 s.

- (2) for example: if it is 08:00, set 0 as number of days to turn the instrument on automatically at 20:00 of the same day; if it is 20:00, set 1 as number of days to turn the instrument on automatically at 08:00 of the following day (you can set the number of days between 0 and 6).

5.2 Activating function Programmed starting

- make sure the instrument is turned on
- press and 1 s: this will turn the instrument off.

At the time you have set with the procedure indicated in paragraph 5.1, the instrument will automatically start working; to turn the instrument on automatically also the following days, repeat the procedure.

The alarm "Real time clock error" interrupts the function.

5.3 Interrupting function Programmed starting

- make sure the instrument is turned off
- press and 1 s.

6 COOKING TIMER

6.1 Setting the cooking timer

To gain access the procedure:

- make sure the instrument is turned on
- press the instrument will show "tin" in the display at the top and the cooking timer value (hours:minutes) in the display at the bottom (the left part of the display will flash).

- press if it is lit, the output for regulation will be turned on

- press if it is lit, the quantity showed in the display at the top will be the working setpoint value

- press if it is lit, the quantity showed in the display at the top will be the chamber temperature

- press if it flashes when the instrument is turned on, the count of the cooking timer will be running

- press if it flashes when the instrument is turned off, function Programmed starting will be active (only RK 801X)

- press or in 15 s (3) (4).

- press during the modification of the hours, then ...

- press or in 15 s (3) (4).

- To quit the procedure:
- do not operate 15 s (4).

- (3) you can set the cooking timer between 00:00 and 24:00 h:min
- (4) you can modify the cooking timer value also if the count is running; if you set 00:00, the instrument will interrupt the function and the buzzer will utter an intermittent beep 3 s.

6.2 Activating the cooking timer

- press during the modification of the minutes.

- make sure the instrument is turned on
- press and 1 s.

6.3 Interrupting the cooking timer

- make sure the instrument is turned on
- press and 1 s.

7 WORKING SETPOINT (WORKING TEMPERATURE)

7.1 Setting the working setpoint

- make sure the instrument is turned on
- press
- press or in 15 s (look at parameters r1 and r2 as well) (5).

- To quit the procedure:
- do not operate 15 s.

- (5) during the modification the instrument will show "SEt" in the display at the bottom.

8 STEAM INJECTION (IF PRESENT)

8.1 Preliminary information

If parameter t0 has value 0, pressing button the instrument will turn the injector on the time you will have set with parameter t2 or as long as you will keep pressed the button; parameter t1 will set the minimum time between two injections in succession.

If parameter t0 has value 1, the instrument will automatically turn the injector on the time you will have set with parameter t2 and will turn the injector off the time you will have set with parameter t1; injection must have been enabled pressing button .

8.2 Setting parameter t2

- press during the modification of the working setpoint, then ...

- press or in 15 s (6).

- To quit the procedure:
- do not operate 15 s.

- (6) during the modification the instrument will show "toN" in the display at the bottom; you can set parameter t2 between 1 and 250 ds.

8.2 Setting parameter t1

- press during the modification of parameter t2, then ...

- press or in 15 s (7).

- To quit the procedure:
- press during the modification of parameter t1 or do not operate 15 s.

- (7) during the modification the instrument will show "tOFF" in the display at the bottom; you can set parameter t1 between 0 and 250 s.

9 CONFIGURATION PARAMETERS

9.1 Setting configuration parameters

To gain access the procedure:

- make sure the instrument is turned off
- press and 4 s: the instrument will show the instrument code in the display at the top to select "PA"

- press or in 15 s to set "19" (8)

- press and 4 s: the instrument will show "P0"

- To select a parameter:
- press or

- To modify a parameter:
- press

- press or in 15 s
- press

- To quit the procedure:
- press and 4 s or do not operate 60 s.

- (8) during the modification the instrument will show the label of the parameter flashing in the display at the top and the parameter value in the display at the bottom.

10 SIGNALS

10.1 Signals

LED	MEANING
out	LED regulator if it is lit, the output for regulation will be turned on
set	LED set if it is lit, the quantity showed in the display at the top will be the working setpoint value if it is switched off, the quantity showed in the display at the top will be the chamber temperature
	LED timer if it flashes when the instrument is turned on, the count of the cooking timer will be running if it flashes when the instrument is turned off, function Programmed starting will be active (only RK 801X)
LED in the bottom right-hand corner	LED steam injection if it is lit: <ul style="list-style-type: none"> the steam injection will be running (if present and if parameter t0 has value 0) the steam injection will have been enabled (if present and if parameter t0 has value 1)
°C	LED Celsius degree if it is lit, the unit of measure of the quantity showed in the display at the top will be Celsius degree
°F	LED Fahrenheit degree if it is lit, the unit of measure of the quantity showed in the display at the top will be Fahrenheit degree
	LED multifunction if it is lit: <ul style="list-style-type: none"> the airhole will be turned on by hand (if present) the chamber light will be turned on (if present) if it flashes, the airhole will automatically be turned on (if present)

	LED ON STAND-BY if it is lit, the instrument will be turned off
--	--

INDICAT.	MEANING
00:01	If it appears in the display at the bottom and the buzzer utters an intermittent beep, there will be 10 s to go before the end of the cooking timer
00:00	If it flashes in the display at the bottom and the buzzer utters an intermittent beep, the count of the cooking timer will be finished

11 ALARMS

11.1 Alarms

CODE	MEANING
AL	Chamber temperature outside safety limits Remedies: <ul style="list-style-type: none"> look at parameters A1 and A4 Effects: <ul style="list-style-type: none"> the alarm output will be activated (if present) the output for acoustic signalings will be activated (if present)

Alarm code flashes alternated with a quantity.

12 INTERNAL DIAGNOSTICS

12.1 Internal diagnostics

CODE	MEANING
SEL	Configuration error Remedies: <ul style="list-style-type: none"> restore default configuration parameters Effects: <ul style="list-style-type: none"> all the outputs will be turned off
PF1	Chamber probe error Remedies: <ul style="list-style-type: none"> look at parameter P0 check the integrity of the probe check the connection instrument probe check the temperature close to the probe Effects: <ul style="list-style-type: none"> the output for regulation will be turned off the output for acoustic signalings will be activated (if present)

14 SETPOINT AND CONFIGURATION PARAMETERS

14.1 Setpoint

	MIN.	MAX.	U.M.	DEF.	SETPOINT
r1	r2	25/50	°C/°F (9)	150	working setpoint

14.2 Configuration parameters

PARAM.	MIN.	MAX.	U.M.	DEF.	MEASURE INPUTS
P0	0	1	---	0	kind of probe (0 = Tc "J", 1 = Tc "K")

